Title	Teknipure C of C			
Document Number	FM-8.2.5.1			
Revision Date	February 20, 2022			

Certificate of Conformance

Wiper Code: TC1PS2-99

LOT No: 527306-110923, 527306-140923,

	s Weight (g STM D3776		Thickness (mm)Orbital Shake Test for ParticlesASTM D1777-96IEST-RP-CC004.3 Section 6.1.4 (x10 ⁶ particle/m ²)		IEST-RP- Sectio	iber/Particles Test For Sorbency Capacity & Rate ST-RP-CC004.3 IEST-RP-C004.3 Section 8.1 Section 6.2.2 (particle/m ²)											
Average	Lower SPEC Limit	Upper SPEC Limit	Thick- ness Average	Thick- ness Lower	Thick- ness Upper	>0.50 µm Average	>0.50 μm SPEC	>5.0 μm Average	>5.0 μm SPEC	>100 µm Average	>100 μm SPEC	Extrinsic Capacity (ml/m ²) Average	Extrinsic Capacity SPEC	Intrinsic Capacity (ml/g) Average	Intrinsic Capacity SPEC	Sorptive Rate (Sec.) Average	Sorptive Rate (Sec.) SPEC
120	115	124	0.46	0.40	0.48	8.5	<10	1.2	<2.0	449	<600	434	>400	2.7	>2.3	<1	<1

	(g/	rt Term Ext m²) .3 Section 7			Specific Extractable lons (ug/g or ppm) IEST-RP-C004.3 Section 7.2.2.1B									
DI Water Average	DI Water SPEC	IPA Average	IPA SPEC	Sodium (Na⁺) Average	Sodium (Na⁺) SPEC	(Na ⁺) (K ⁺)		Calcium (Ca ²⁺) Average	Calcium (Ca ²⁺) SPEC	Chloride (CI ⁻) Average	Chloride (Cl ⁻) SPEC	Magnesium (Mg ²⁺) Average	Magnesium (Mg²⁺) SPEC	
0.013	< 0.02	0.061	<0.080	0.21	< 0.30	0.014	<0.20	0.27	<0.40	0.10	<0.20	0.22	<0.30	

Product Cleanliness Specifications	Tested By: (Production Technician)				
The raw materials used in the production of this finished product are compliant with IEST-RP-CC004.3 and in accordance with Teknipure specifications.	张志馨				
	Inspected By: (Quality)				
Product Packaging Specifications					
The packaging materials, method & environment of this finished product are in accordance with	王雪艳				
Teknipure specifications.	Name (printed):				
Shelf Life	王雪艳				
	Date of Manufacture:				
Sterile product is marked with an expiration date three (3) years from date of manufacture.					
	<u>Sep.11/ Sep.14.2023</u>				



Innovative Contamination Solutions