



Innovative
Contamination
Solutions

Document Title:	Teknisat C of C
Document Number:	FM-8.2.4-2
Approved By:	Ben Frazier
Revision Date:	February 10, 2017

Certificate of Conformance

Wiper Code: TC2PL1-99G

Lot no 526270

I. Physical Property Variables

Property	Target	Lower	Upper	Units	Test Method
Basis Weight	140	135	145	g/m ²	Weight Balance
Thickness	0.45	0.40	0.50	mm	Caliper

II. Test for Particles ($\times 10^6$ particle/m²) - Orbital Shake Test (*IEST-RP-CC004.3 Section 6.1.4*)

Particle size	Result (Average of 3 trials)	Specs	JUDGMENT
> 0.50 μ m	3.1	< 11.0	PASS
> 5.0 μ m	0.5	< 0.8	PASS

III. Test for Fiber /Particles >100 μ m ($\times 10^6$ particle/m²) (*IEST-RP-CC004.3 Section 6.2.2*)

Particle size	Result (Average of 3 trials)	Specs	JUDGMENT
> 100 μ m	135	< 400	PASS

IV. Test for Sorbency Capacity and Rate – (*IEST-RP-C004.3 Section 8.1*)

Note: Unless specified the Default Test Solution is DI Water

Absorbency	Result (Ave of 3 trials)	Specs	JUDGMENT
Extrinsic Capacity (mL/m ²)	416	> 410	PASS
Intrinsic Capacity (mL/g)	2.95	> 2.60	PASS
Sorptive Rate(S)	< 1	< 1	PASS

V. Test for Extractable Matters (g/m²) - Short Term Extraction (*IEST-RP-CC004.3 Section 7.1.2*)

Test Solution	Result (Average of 3 trials)	Specs	JUDGMENT
DI Water	0.006	< 0.03	PASS
IPA	0.011	< 0.08	PASS

VI. Test for Specific Extractable Ions (ug/g or ppm)

Standard Extraction Method (*IEST-RP-C004.3 Section 7.2.2.1B*)

Extractable Ion	Result(Ave of 3 trials)	Specs	JUDGMENT
Sodium (Na ⁺)	0.15	< 0.20	PASS
Potassium (K ⁺)	0.08	< 0.10	PASS
Calcium(Ca ²⁺)	0.13	< 0.20	PASS
Chloride (Cl ⁻)	0.07	< 0.10	PASS
Magnesium (Mg ²⁺)	0.08	< 0.10	PASS

Tested By: Bill
 Inspected By: 李伟光
 Name (Printed): Bill
 Date of manufacture Jan. 3, 2019

