



Innovative  
Contamination  
Solutions

Document Title:	Teknisat C of C
Document Number:	FM-8.2.4-2
Approved By:	Eric Hilkmann
Revision Date:	February 10, 2017

## Certificate of Conformance

Wiper Code: TC2PL1-99G

Lot no 526693

### I. Physical Property Variables

Property	Target	Lower	Upper	Units	Test Method
Basis Weight	140	135	145	g/m <sup>2</sup>	Weight Balance
Thickness	0.45	0.40	0.50	mm	Caliper

### II. Test for Particles (x 10<sup>6</sup>particle/m<sup>2</sup>) - Orbital Shake Test (IEST-RP-CC004.3 Section 6.1.4)

Particle size	Result (Average of 3 trials)	Specs	JUDGMENT
> 0.50 um	3.1	< 11.0	PASS
> 5.0 um	0.5	< 0.8	PASS

### III. Test for Fiber /Particles >100um (x10<sup>6</sup>particle/m<sup>2</sup>) (IEST-RP-CC004.3 Section 6.2.2)

Particle size	Result (Average of 3 trials)	Specs	JUDGMENT
> 100 um	114	< 400	PASS

### IV. Test for Sorbency Capacity and Rate – (IEST-RP-C004.3 Section 8.1)

Note: Unless specified the Default Test Solution is DI Water

Absorbency	Result (Ave of 3 trials)	Specs	JUDGMENT
Extrinsic Capacity (ml/m <sup>2</sup> )	418	> 410	PASS
Intrinsic Capacity (ml/g)	2.95	>2.60	PASS
Sorptive Rate(S)	< 1	< 1	PASS

### V. Test for Extractable Matters (g/m<sup>2</sup>) - Short Term Extraction (IEST-RP-CC004.3 Section 7.1.2)

Test Solution	Result (Average of 3 trials)	Specs	JUDGMENT
DI Water	0.015	< 0.03	PASS
IPA	0.022	< 0.08	PASS

### VI. Test for Specific Extractable Ions (ug/g or ppm) Standard Extraction Method (IEST-RP-C004.3 Section 7.2.2.1B)

Extractable Ion	Result(Ave of 3 trials)	Specs	JUDGMENT
Sodium (Na <sup>+</sup> )	0.16	< 0.20	PASS
Potassium (K <sup>+</sup> )	0.07	< 0.10	PASS
Calcium(Ca <sup>2+</sup> )	0.15	< 0.20	PASS
Chloride (Cl <sup>-</sup> )	0.08	< 0.10	PASS
Magnesium (Mg <sup>2+</sup> )	0.09	<0.10	PASS

Tested By: Bill  
 Inspected By: 张伟光  
 Name (Printed): Bill  
 Date of manufacture Aug. 28, 2020

