



Innovative
Contamination
Solutions

Document Title:

Teknisat C of C

Document Number:

FM-8.2.4-2

Approved By:

Eric Hillman

Revision Date:

February 10, 2017

Certificate of Conformance

Wiper Code: TC2PL1-99G

Lot no 526757

I. Physical Property Variables

Property	Target	Lower	Upper	Units	Test Method
Basis Weight	140	135	145	g/m ²	Weight Balance
Thickness	0.45	0.40	0.50	mm	Caliper

II. Test for Particles (x 10⁶particle/m²) - Orbital Shake Test

(*IEST-RP-CC004.3 Section 6.1.4*)

Particle size	Result (Average of 3 trials)	Specs	JUDGMENT
> 0.50 um	3.6	< 11.0	PASS
> 5.0 um	0.7	< 0.8	PASS

III. Test for Fiber/Particles >100um (x10⁶particle/m²)

(*IEST-RP-CC004.3 Section 6.2.2*)

Particle size	Result (Average of 3 trials)	Specs	JUDGMENT
> 100 um	118	< 400	PASS

IV. Test for Sorbency Capacity and Rate – (*IEST-RP-C004.3 Section 8.1*)

Note: Unless specified the Default Test Solution is DI Water

Absorbency	Result (Ave of 3 trials)	Specs	JUDGMENT
Extrinsic Capacity (ml/m ²)	419	> 410	PASS
Intrinsic Capacity (ml/g)	2.94	>2.60	PASS
Sorptive Rate(S)	< 1	< 1	PASS

V. Test for Extractable Matters (g/m²) - Short Term Extraction

(*IEST-RP-CC004.3 Section 7.1.2*)

Test Solution	Result (Average of 3 trials)	Specs	JUDGMENT
DI Water	0.017	< 0.03	PASS
IPA	0.024	< 0.08	PASS

VI. Test for Specific Extractable Ions (ug/g or ppm)

Standard Extraction Method (IEST-RP-C004.3 Section 7.2.2.1B)

Extractable Ion	Result(Ave of 3 trials)	Specs	JUDGMENT
Sodium (Na ⁺)	0.15	< 0.20	PASS
Potassium (K ⁺)	0.08	< 0.10	PASS
Calcium(Ca ²⁺)	0.14	< 0.20	PASS
Chloride (Cl ⁻)	0.07	< 0.10	PASS
Magnesium (Mg ²⁺)	0.08	<0.10	PASS

Tested By: Bill

Inspected By: 吕伟光

Name (Printed): Bill

Date of manufacture Oct. 12, 2020

