

Title	Teknipse C of C
Document Number	FM-8.2.5.1
Revision Date	February 20, 2022

Certificate of Conformance

Wiper Code: TC2PL2.99G

LOT No: 270722

Basis Weight (g/m ²) ASTM D3776		Thickness (mm) ASTM D1777-96		Orbital Shake Test for Particles IEST-RP-CC004.3 Section 6.1.4 ($\times 10^6$ particle/m ²)			Fiber/Particles IEST-RP-CC004.3 Section 6.2.2 (particle/m ²)			Test For Sorbency Capacity & Rate IEST-RP-CC004.3 Section 8.1							
Average	Lower SPEC Limit	Upper SPEC Limit	Thickness Average	Thickness Upper	Thickness Lower	>0.50 μm Average	>5.0 μm μm SPEC	>5.0 μm μm SPEC	>100 μm Average	>100 μm SPEC	Intrinsic Capacity (ml/g) Average	Intrinsic Capacity (ml/g) SPEC	Intrinsic Capacity (ml/g) SPEC	Sorptive Rate (Sec.) SPEC			
121	115	125	0.42	0.35	0.45	3.7	< 5.0	0.3	< 0.5	103	<4 00	335	>300	2.63	>2.5	< 1	< 1

Extractables - Short Term Extraction (g/m ²) IEST-RP-CC004.3 Section 7.1.2										Specific Extractable Ions (ug/g or ppm) IEST-RP-CC004.3 Section 7.2.2.1B				
DI Water Average	DI Water SPEC	IPA Average	IPA SPEC	Sodium (Na ⁺) Average	Sodium (Na ⁺) SPEC	Potassium (K ⁺) Average	Potassium (K ⁺) SPEC	Calcium (Ca ²⁺) Average	Calcium (Ca ²⁺) SPEC	Chloride (Cl ⁻) Average	Chloride (Cl ⁻) SPEC	Magnesium (Mg ²⁺) Average	Magnesium (Mg ²⁺) SPEC	
0.0033	< 0.01	0.019	< 0.04	0.10	< 0.20	0.09	< 0.10	0.07	< 0.20	0.07	< 0.10	0.08	< 0.10	

Product Cleanliness Specifications

The raw materials used in the production of this finished product are compliant with IEST-RP-CC004.3 and in accordance with Teknipse specifications.

Product Packaging Specifications

The packaging materials, method & environment of this finished product are in accordance with Teknipse specifications.

Shelf Life

Sterile product is marked with an expiration date three (3) years from date of manufacture.

Tested By: (Production Technician) B:11

Inspected By: (Quality) B:16 7/27/22

Name (printed): B:11

Date of Manufacture: July 27, 2022



Innovative
Contamination
Solutions

Toll Free: 844.309.2376 Quality@Teknipse.com www.Teknipse.com