Title	Teknipure C of C			
Document Number	FM-8.2.5.1			
Revision Date	February 20, 2022			

Certificate of Conformance

Wiper Code: TC2PU1-12

LOT No: 527181-170123

	Basis Weight (g/m²) ASTM D3776		Thickness (mm) ASTM D1777-96			Orbital Shake Test for Particles IEST-RP-CC004.3 Section 6.1.4 (x10 ⁶ particle/m ²)			Fiber/Particles IEST-RP-CC004.3 Section 6.2.2 (particle/m ²)		Test For Sorbency Capacity & Rate IEST-RP-C004.3 Section 8.1						
Average	Lower SPEC Limit	Upper SPEC Limit	Thick- ness Average	Thick- ness Lower	Thick- ness Upper	>0.50 µm Average	>0.50 μm SPEC	>5.0 μm Average	>5.0 μm SPEC	>100 μm Average	>100 μm SPEC	Extrinsic Capacity (ml/m ²) Average	Extrinsic Capacity SPEC	Intrinsic Capacity (ml/g) Average	Intrinsic Capacity SPEC	Sorptive Rate (Sec.) Average	Sorptive Rate (Sec.) SPEC
140	135	142	0.45	0.43	0.48	4.05	<5.0	0.12	<0.3	328	<400	421	>400	3.1	>2.8	<1	<1

IEST-F	Extractio	bles - Shor on (g/m²) .3 Section 7		Specific Extractable lons (ug/g or ppm) IEST-RP-C004.3 Section 7.2.2.1B									
DI Water Average	DI Water SPEC	IPA Average	IPA SPEC	Sodium (Na⁺) Average	Sodium (Na⁺) SPEC	Potassium (K⁺) Average	Potassium (K⁺) SPEC	Calcium (Ca²⁺) Average	Calcium (Ca ²⁺) SPEC	Chloride (Cl ⁻) Average	Chloride (Cl ⁻) SPEC	Magnesium (Mg ²⁺) Average	Magnesium (Mg ²⁺) SPEC
0.007	<0.01	0.030	< 0.04	0.010	<0.02	0.006	<0.01	0.012	<0.02	0.010	<0.02	0.012	<0.02

Product Cleanliness Specifications	Tested By: (Production Technician)				
The raw materials used in the production of this finished product are compliant with IEST-RP-CC004.3 and in accordance with Teknipure specifications.	张志奏				
Product Packaging Specifications	Inspected By: (Quality)				
The packaging materials, method & environment of this finished product are in accordance with	王雪艳				
Teknipure specifications.	Name (printed):				
Shelf Life	王雪艳				
Sterile product is marked with an expiration date three (3) years from date of manufacture.	Date of Manufacture:				
	Jan.17.2023				

Innovative Teknipure

Contamination Solutions