



Innovative
Contamination
Solutions

Document:	Teknisat C of C
Document:	FM-8.2
Approved:	Ben
Revision:	February 10, 2017

Certificate of Conformance

Wiper Code: TC2PU1-99

Lot no:526625

I. Physical Property Variables

Property	Target	Lower	Upper	Units	Test Method
Basis Weight	140	135	145	g/m ²	Weight Balance
Thickness	0.50	0.45	0.55	mm	Caliper

II. Test for Particles ($\times 10^6$ particle/m²) - Orbital Shake Test (*IEST-RP-CC004.3 Section 6.1.4*)

Particle size	Result (Average of 3 trials)	Specs	JUDGMENT
> 0.50 μ m	3.4	< 5.0	PASS
> 5.0 μ m	0.13	< 0.5	PASS

III. Test for Fiber /Particles >100 μ m ($\times 10^6$ particle/m²) (*IEST-RP-CC004.3 Section 6.2.2*)

Particle size	Result (Average of 3 trials)	Specs	JUDGMENT
> 100 μ m	142	< 400	PASS

IV. Test for Sorbency Capacity and Rate - (*IEST-RP-C004.3 Section 8.1*)

Note: Unless specified the Default Test Solution is DI Water

Absorbency	Result (Ave of 3 trials)	Specs	JUDGMENT
Extrinsic Capacity (ml/m ²)	418	> 410	PASS
Intrinsic Capacity (ml/g)	2.98	>2.6	PASS
Sorptive Rate(S)	< 1	< 1	PASS

V. Test for Extractable Matters (g/m²) - Short Term Extraction (*IEST-RP-CC004.3 Section 7.1.2*)

Test Solution	Result (Average of 3 trials)	Specs	JUDGMENT
DI Water	0.0033	< 0.01	PASS
IPA	0.017	< 0.04	PASS

VI. Test for Specific Extractable Ions (μ g/g or ppm)

Standard Extraction Method (IEST-RP-C004.3 Section 7.2.2.1B)

Extractable Ion	Result(Ave of 3 trials)	Specs	JUDGMENT
Sodium (Na ⁺)	0.11	< 0.20	PASS
Potassium (K ⁺)	0.08	< 0.10	PASS
Calcium(Ca ²⁺)	0.17	< 0.20	PASS
Chloride (Cl)	0.09	< 0.10	PASS
Magnesium (Mg ²⁺)	0.08	<0.10	PASS

Tested By: Bill

Inspected By: 张伟光

Name (Printed): Bill

Date of manufacture May 14, 2020

