Title	Teknipure C of C			
Document Number	FM-8.2.5.1			
Revision Date	February 20, 2022			

Certificate of Conformance

Wiper Code: TC3PB1-12

LOT No: 527272-100723, 527272-130723

	s Weight (g ASTM D377			ickness (mi TM D1777-	•	Orbital Shake Test for Particles IEST-RP-CC004.3 Section 6.1.4 (x10 ⁶ particle/m ²)			IEST-RP	Particles -CC004.3 n 6.2.2 :le/m ²)		Test For Sorbency Capacity & Rate IEST-RP-C004.3 Section 8.1					
Average	Lower SPEC Limit	Upper SPEC Limit	Thick- ness Average	Thick- ness Lower	Thick- ness Upper	>0.50 µm Average	>0.50 μm SPEC	>5.0 μm Average	>5.0 μm SPEC	>100 µm Average	>100 μm SPEC	Extrinsic Capacity (ml/m ²) Average	Extrinsic Capacity SPEC	Intrinsic Capacity (ml/g) Average	Intrinsic Capacity SPEC	Sorptive Rate (Sec.) Average	Sorptive Rate (Sec.) SPEC
140	135	142	0.45	0.43	0.48	4.3	<5.0	0.14	<0.3	341	<400	423	>400	3.1	>2.8	<1	<1

	(g/	ort Term Ext m ²) I.3 Section 1			Specific Extractable Ions (ug/g or ppm) IEST-RP-C004.3 Section 7.2.2.1B									
DI Water Average	DI Water SPEC	IPA Average	IPA SPEC	Sodium (Na⁺) Average	Sodium (Na⁺) SPEC	(Na ⁺) (K ⁺)		Potassium Calcium (K ⁺) SPEC (Ca ²⁺) Average		Chloride (Cl ⁻) Average	Chloride (Cl ⁻) SPEC	Magnesium (Mg ²⁺) Average	Magnesium (Mg²⁺) SPEC	
0.007	<0.01	0.021	<0.04	0.012	< 0.02	0.006	<0.01	0.008	<0.02	0.010	< 0.02	0.014	<0.02	

Product Cleanliness Specifications	ested By: (Production Technician)
The raw materials used in the production of this finished product are compliant with IEST-RP-CC004.3 and in accordance with Teknipure specifications.	张志奏
Product Packaging Specifications	Inspected By: (Quality)
Froduct Fackaging Specifications	
The packaging materials, method & environment of this finished product are in accordance with	王雪艳
Teknipure specifications.	Name (printed):
Shelf Life	王雪艳
	Date of Manufacture:
Sterile product is marked with an expiration date three (3) years from date of manufacture.	
	July.10/ July.13.2023

Innovative Teknipure

Contamination Solutions