Revision Date Document Number FM-8.2.5.1 February 20, 2022

Certificate of Conformance

Wiper Code: TC3PB1S-12

LOT No: 527271-100823

Basis A	Basis Weight (g/m² ASTM D3776	/m²) 6	AS	Thickness (mm) ASTM D1777-96	.96 m)	Orbit IEST-	al Shake T RP-CC004 (x10 ⁶ par	Orbital Shake Test for Particles IEST-RP-CC004.3 Section 6.1.4 (x10 ⁶ particle/m²)	1.4	Fiber/Particles IEST-RP-CC004 Section 6.2.2 (particle/m²)	Fiber/Particles EST-RP-CC004.3 Section 6.2.2 (particle/m²)		Test	t For Sorbency Capacity & F IEST-RP-C004.3 Section 8.1	νς: 3.5	ection 8.	Test For Sorbency Capacity & Rate IEST-RP-C004.3 Section 8.1
Average	Lower SPEC	Upper SPEC	Thick- ness	Thick- ness	Thick- ness	μm >0.50	>0.50	>5.0 µm Average	μm 9.5<	>100 µm	>100 µm	Extrinsic Capacity	Extrinsic Capacity	Intrin	sic	sic Intrinsic city Capacity	Intrinsic Sorptive Capacity Rate
	Limit	Limit	Average	Lower	Upper	Average	SPEC		SPEC	Average	SPEC	(ml/m²) Average	SPEC	(ml/g) Average) ge		
140	135	145	0.45	0.40	0.50	3.1	< 5.0	0.4	< 0.5	108	<4 00	418	> 400	2.99		9 >2.8	

0.0047	DI Water Average	Extracta
< 0.01	DI Water SPEC	oles - Sho (g/ RP-CC004
0.0047 < 0.01 0.017 < 0.04 0.16	IPA Average	Extractables - Short Term Extraction (g/m²) IEST-RP-CC004.3 Section 7.1.2
< 0.04	IPA SPEC	raction 7.1.2
0.16	Sodium (Na ⁺) Average	
< 0.20	Sodium (Na ⁺) SPEC	
0.08	Potassium (K+) Average	
< 0.10	Potassium (K¹) SPEC	Specific IEST
0.07	Calcium (Ca ²⁺) Average	-RP-C004.3
< 0.20	n Calcium (Ca²+) e SPEC	Specific Extractable lons (ug/g or ppm) IEST-RP-C004.3 Section 7.2.2.1B
0.08	Chloride (Cl ⁻) Average	or ppm) 2.2.1B
< 0.10	Chloride (CI ⁻) SPEC	
0.07	Magnesium (Mg²+) Average	
< 0.10	Magnesium (Mg²+) SPEC	

Product Cleanliness Specifications

The raw materials used in the production of this finished product are compliant with IEST-RP-CC004.3 and in accordance with Teknipure specifications.

Product Packaging Specifications

Teknipure specifications. The packaging materials, method & environment of this finished product are in accordance with

Shelf Life

Sterile product is marked with an expiration date three (3) years from date of manufacture.

Name (printed): 811 Inspected By: (Quality) Date of Manufacture: Tested By: (Production Technician) Aug. 10, 223 May B::





Toll Free: 844.309.2376 🔄 Quality@Teknipure.com 🌐 www.Teknipure.com



