



Document Title:	Teknipure C of C
Document Number:	FM-8.2.4-7
Approved By:	Eric Hillman
Revision Date:	Eight-4-2021

Certificate of Conformance

Wiper Code:TS-MW-3

Lot No:527140

I. Physical Property Variables

Property	Target	Lower	Upper	Units	Test Method
Head Length	12	11	12	mm	Ruler
Handle Length	60	59	60	mm	Ruler
Total Length	70	69	70	mm	Ruler

II. Test for Particles (particles/swab)- Orbital Shake Test

Particle size	Result (Average of 3 trials)	Specs	JUDGMENT
> 0.5 um	3310	< 4000	PASS

III. Test for Extractable Matters (mg/swab) • Short Term Extraction

Test Solution	Result (Average of 3 trials)	Specs	JUDGMENT
DI Water	0.0613	< 1.00	PASS
DI IPA	0.1400	< 1.00	PASS

IV. Test for Specific Extractable Ions (ug/swab) - Standard Extraction Method

Extractable Ions	Result (Average of 3 trials)	Specs	JUDGMENT
Chloride (Cl)	ND	<2.00	PASS
Sulphate (SO ₄ ²⁻)	0.0989	< 2.00	PASS

The packaging materials,method & environment of this finished product are in accordance with Teknipure specifications.

Tested By: Liang Jia Tu
Laboratory Manager: Xu Jing
Date of Manufacture: 10.08.2022