



Innovative  
Contamination  
Solutions

### Certificate of Conformance

Document Title	Teknstat C. of C
Document Number	FM-8.2.4-2
Approved By	Ben Frazier
Revision Date	February 10, 2017

**Wiper Code:** TS1MPI70-911

**Lot No:** 415476

#### **I. Component Traceability**

Component	PN#	LOT #
Wiper	TZ1MPB-911	526365
Fluid	2-Propanol; VLSI/USP Grade	A03092009W

#### **II. Saturation & Fluid Specifications**

WFI Resistivity	WFI TOC (ppb)	Total Fluid Volume/Bag (ml)	IPA Fluid Volume/Bag (ml)	IPA Fluid %	UPW Fluid Volume/Bag (ml)	UPW Fluid %
>17.8 (M. Ohm)	<5	410	287	70	123	30
18.51	2.89					

#### **III. Product Cleanliness Specifications**

The raw materials used in the production of this finished product are compliant per IEST-RP-CC004.3 and in accordance with Teknipure upper limit specifications. Fluids meet specifications & traceability of USP and SEMI established criteria with lot# CofA on file. WFI has TDS, TOC, & Endoxin Removal <1/EU

#### **IV. Product Packaging Specifications**

The packaging materials, method & environment of this finished product are in accordance with Teknipure specifications.

#### **V. Expiration Date**

Expiration date is two (2) years from date of manufacture.

Manufactured By: Production Technician: Wainner m.

Inspected By: Gary Barrow

Name (printed): GARY BARROW

Date of Manufacture: 03/26/2020