



Innovative
Contamination
Solutions

Certificate of Conformance

| | |
|-----------------|------------------|
| Document Title | Teknistat C of C |
| Document Number | FM-8.2.4-2 |
| Approved By | Eric Hillkman |
| Revision Date | July 29, 2020 |

Wiper Code: TS1MPI70-911

Lot No: 415682

I. Component Traceability

| Component | PN# | LOT # |
|-----------|----------------------------|------------|
| Wiper | TZ1MPB-911 | 526626 |
| Fluid | 2-Propanol; VLSI/USP Grade | DA20885133 |

II. Saturation & Fluid Specifications

| UPW Resistivity | UPW TOC (ppb) | Total Fluid Volume/Bag (ml) | IPA Fluid Volume/Bag (ml) | IPA Fluid % | UPW Fluid Volume/Bag (ml) | UPW Fluid % |
|-----------------|---------------|-----------------------------|---------------------------|-------------|---------------------------|-------------|
| >17.8 (M. Ohm) | <5 | 410 | 287 | 70 | 123 | 30 |
| 18.55 | 4.99 | | | | | |

III. Product Cleanliness Specifications

The raw materials used in the production of this finished product are compliant per IEST-RP-CC004.3 and in accordance with Teknipure upper limit specifications and fluids meet specifications of USP and SEMI established criteria.

IV. Product Packaging Specifications

The packaging materials, method & environment of this finished product are in accordance with Teknipure specifications.

V. Expiration Date

Expiration date is two (2) years from date of manufacture.

Manufactured By: Production Technician: [Signature]

Inspected By: [Signature]

Name (printed): ERIC HILLKMAN

Date of Manufacture: 09/03/2020