



Innovative
Contamination
Solutions

Certificate of Conformance

| | |
|-----------------|-------------------|
| Document Title | Teknistat C of C |
| Document Number | FM-8.2.4-2 |
| Approved By | Ben Frazier |
| Revision Date | February 10, 2017 |

Wiper Code: TS2PL2I70R-99 (F)

Lot No: 415570

I. Component Traceability

| Component | PN# | LOT # |
|-----------|----------------------------|------------|
| Wiper | TC2PL2-99 | 526365 |
| Fluid | 2-Propanol; VLSI/USP Grade | A06232009B |

II. Saturation & Fluid Specifications

| WFI Resistivity | WFI TOC (ppb) | Total Fluid Volume/Bag (ml) | IPA Fluid Volume/Bag (ml) | IPA Fluid % | UPW Fluid Volume/Bag (ml) | UPW Fluid % |
|-----------------|---------------|-----------------------------|---------------------------|-------------|---------------------------|-------------|
| >17.8 (M. Ohm) | <5 | 800 | 560 | 70 | 240 | 30 |
| 18.57 | 4.99 | | | | | |

III. Product Cleanliness Specifications

The raw materials used in the production of this finished product are compliant per IEST-RP-CC004.3 and in accordance with Teknipure upper limit specifications. Fluids meet specifications & traceability of USP and SEMI established criteria with lot# CofA on file. WFI has TDS, TOC, & Endoxin Removal <1/EU

IV. Product Packaging Specifications

The packaging materials, method & environment of this finished product are in accordance with Teknipure specifications.

V. Expiration Date

Expiration date is two (2) years from date of manufacture.

Manufactured By: Production Technician: 

Inspected By: 

Name (printed): Eric Hilkmann

Date of Manufacture: 07/08/2020