

Certificate of Conformance

Document TitleTeknistat C of CDocument NumberFM-8.2.4-2Approved ByBen FrazierRevision DateFebruary 10, 2017

Wiper Code: TS2PUI70R-44

Lot No: 415615

I. Component Traceability

| Component | PN# | LOT# | |
|-----------|----------------------------|------------|--|
| Wiper | TC2PU1-44 | 526581 | |
| Fluid | 2-Propanol; VLSI/USP Grade | A04222010Y | |

II. Saturation & Fluid Specifications

| WFI Resistivity | WFI TOC (ppb) | Total Fluid Volume/Bag (ml) | IPA Fluid Volume/Bag (ml) | IPA Fluid % | UPW Fluid Volume/Bag (ml) | UPW Fluid % |
|--------------------|------------------|-----------------------------------|---------------------------------|----------------|---------------------------------|----------------|
| >17.8 (M. Ohm) | <5 | 250 | 175 | 70 | 75 | 30 |
| 18.54 | 4.32 | | | | | |

III. Product Cleanliness Specifications

The raw materials used in the production of this finished product are compliant per IEST-RP-CC004.3 and in accordance with Teknipure upper limit specifications. Fluids meet specifications & traceability of USP and SEMI established criteria with lot# CofA on file. WFI has TDS, TOC, & Endoxin Removal <1/EU

IV. Product Packaging Specifications

The packaging materials, method & environment of this finished product are in accordance with Teknipure specifications.

V. Expiration Date

| Expiration date is two (2) years from date of manufacture. |
|--|
| Manufactured By: Broduction Technician: |
| Inspected By: TOPUS CERREA |
| Name (printed): Or one Cornea |
| Date of Manufacture: 05/20/2020 |