



Innovative
Contamination
Solutions

Certificate of Conformance

Document Title	Teknistat C of C
Document Number	FM-8.2.4-2
Approved By	Ben Frazier
Revision Date	February 10, 2017

Wiper Code: TS3PBI70S-12

Lot No: 415025

I. Component Traceability

Component	PN#	LOT #
Wiper	TC3PB1S-12	526322
Fluid	2-Propanol; VLSI/USP Grade	A08051908N

II. Saturation & Fluid Specifications

WFI Resistivity	WFI TOC (ppb)	Total Fluid Volume/Bag (ml)	IPA Fluid Volume/Bag (ml)	IPA Fluid %	UPW Fluid Volume/Bag (ml)	UPW Fluid %
>17.8 (M. Ohm)	<5	1800	1260	70	540	30
18.47	3.37					

III. Product Cleanliness Specifications

The raw materials used in the production of this finished product are compliant per IEST-RP-CC004.3 and in accordance with Teknipure upper limit specifications. Fluids meet specifications & traceability of USP and SEMI established criteria with lot# CofA on file. WFI has TDS, TOC, & Endoxin Removal <1/EU

IV. Product Packaging Specifications

The packaging materials, method & environment of this finished product are in accordance with Teknipure specifications.

V. Expiration Date

Expiration date is two (2) years from date of manufacture.

Manufactured By: Production Technician: Brandy Thomas

Inspected By: Mary Barrow

Name (printed): GARY BARROW

Date of Manufacture: 08/22/2019