Certificate of Conformance

Wiper Code: TZ1MPB-911

LOT No: 527288-010823

Basis Weight (g/m²) ASTM D3776			Thickness (mm) ASTM D1777-96			Orbital Shake Test for Particles IEST-RP-CC004.3 Section 6.1.4 (x10⁶ particle/m²)				Fiber/Particles IEST-RP-CC004.3 Section 6.2.2 (particle/m²)		Test For Sorbency Capacity & Rate IEST-RP-CC004.3 Section 8.1					
Lower SPEC Limit	Upper SPEC Limit	Average	Thickness Average	Thickness Lower	Thickness Upper	>0.50 μm Average	>0.50 μm SPEC	>5.0 μm Average	>5.0 μm SPEC	>100 μm Average	>100 μm SPEC	Extrinsic Capacity (ml/m²) Average	Extrinsic Capacity (ml/m²) SPEC	Intrinsic Capacity (ml/g) Average	Intrinsic Capacity (ml/g) SPEC	Sorptive Rate (Sec.) Average	Sorptive Rate (Sec.) SPEC
33	37	35	0.20	0.18	0.22	17.5	< 20.0	1.9	<3.0	183	<1000	NA	NA	NA	NA	NA	NA

Extractables - Short Term Extraction (g/m²) IEST-RP-CC004.3 Section 7.1.2

DI Water Average	DI Water SPEC	IPA Average	IPA SPEC
0.004	<0.005	0.007	< 0.01

Specific Extractable Ions (ug/g or ppm) IEST-RP-CC004.3 Section 7.2.2.1B

Sodium (Na⁺) Average	Sodium (Na⁺) SPEC	Potassium (K⁺) Average	Potassium (K⁺) SPEC	Calcium (Ca²⁺) Average	Calcium (Ca²⁺) SPEC	Chloride (Cl⁻) Average	Chloride (Cl⁻) SPEC	Magnesium (Mg²⁺) Average	Magnesium (Mg²⁺) SPEC
0.35	< 0.50	0.23	< 0.50	0.08	< 0.10	0.07	< 0.10	0.08	< 0.10

Tested By: (Production Technician) _Bill_

Inspected By: (Quality) _MaM_

Name (printed): _Bill_

Date of Manufacture: _Aug, 1, 2023_

Product Cleanliness Specifications

The raw materials used in the production of this finished product are compliant with IEST-RP-CC004.3 and in accordance with Teknipure specifications.

Product Packaging Specifications

The packaging materials, method & environment of this finished product are in accordance with Teknipure specifications.

Shelf Life

Sterile product is marked with an expiration date three (3) years from date of manufacture.

Teknipure | Innovative Contamination Solutions

Toll Free: 844.309.2376 Quality@Teknipure.com www.Teknipure.com