

	Document Title:	Teknipure C of C
	Document Number:	FM-8.2.4-7
	Approved By:	Eric Hillman
	Revision Date:	June-25-2021

Certificate of Conformance

Wiper Code: TZ1PCS1-44

Lot No: 527037

I. Physical Property Variables

Property	Target	Lower	Upper	Units	Test Method
Basis Weight	69	68	69	g/m ²	Weight Balance
Thickness	0.28	0.27	0.28	mm	Caliper

II. Test for Particles (x 106particle/m²)- Orbital Shake Test(IEST-RP-CC004.3 Section 6.1.4)

Particle size	Result (Average of 3 trials)	Specs	JUDGMENT
> 0.50 um	77	< 80	PASS
> 5.0 um	0.19	< 0.26	PASS

III. Test for Fiber /Particles >100um (x106particle/m²)(IEST-RP-CC004.3 Section 6.2.2)

Particle size	Result (Average of 3 trials)	Specs	JUDGMENT
> 100 um	264	< 300	PASS

IV. Test for Sorbency Capacity and Rate- (IEST-RP-CC004.3 Section 8.1)Note: Unless specified the Default Test Solution is DI Water

Absorbency	Result (Average of 3 trials)	Specs	JUDGMENT
Extrinsic Capacity (mL/m ²)	381	> 280	PASS
Intrinsic Capacity (mL/g)	5.72	> 4.0	PASS
Sorptive Rate (s)	0.41	< 1	PASS

V. Test for Extractable Matters (g/m²) • Short Term Extraction (IEST-RP-CC004.3 Section 7.1.2)

Test Solution	Result (Average of 3 trials)	Specs	JUDGMENT
DI Water	0.0447	< 0.070	PASS
IPA	0.0017	< 0.006	PASS

VI. Test for Specific Extractable Ions (ug/g or ppm) - Standard Extraction Method (IEST-RP-CC004.3 Section 7.2.2.18)

Extractable Ions	Result (Average of 3 trials)	Specs	JUDGMENT
Sodium (Na ⁺)	42.18	< 50.0	PASS
Potassium (K ⁺)	22.49	< 30.0	PASS
Calcium (Ca ²⁺)	14.47	< 30.0	PASS
Chloride (Cl ⁻)	21.86	< 30.0	PASS
Magnesium (Mg ²⁺)	6.38	< 50.0	PASS

The packaging materials, method & environment of this finished product are in accordance with Teknipure specifications.

Tested By: Liang Jia EN

Laboratory Manager: chen shi quan

Approved By: Yang guang jue

Date of Manufacture: 2021.12.19