

	Document Title:	Teknipure C of C
	Document Number:	FM-8.2.4-7
	Approved By:	Ben
	Revision Date:	November-26-2018

Certificate of Conformance

Wiper Code: TZ1PCS1-77

Lot No: 526254

I. Physical Property Variables

Property	Target	Lower	Upper	Units	Test Method
Basis Weight	69	66	67	g/m ²	Weight Balance
Thickness	0.28	0.26	0.28	mm	Caliper

II. Test for Particles ($\times 10^6$ particle/m²)- Orbital Shake Test (IEST-RP-CC004.3 Section 6.1.4)

Particle size	Result (Average of 3 trials)	Specs	JUDGMENT
> 0.50 μ m	49	< 80	PASS
> 5.0 μ m	0.14	< 0.26	PASS

III. Test for Fiber /Particles >100 μ m ($\times 10^6$ particle/m²) (IEST-RP-CC004.3 Section 6.2.2)

Particle size	Result (Average of 3 trials)	Specs	JUDGMENT
> 100 μ m	232	< 300	PASS

IV. Test for Sorbency Capacity and Rate- (IEST-RP-C004.3 Section 8.1)

Note: Unless specified the Default Test Solution is 01 Water

Absorbency	Result (Average of 3 trials)	Specs	JUDGMENT
Extrinsic Capacity (mL/m ²)	522	> 280	PASS
Intrinsic Capacity (mL/g)	8.23	> 4.0	PASS
Sorptive Rate (s)	0.5	< 1	PASS

V. Test for Extractable Matters (g/m²) • Short Term Extraction (IEST-RP-CC0.3 Section 7.1.2)

Test Solution	Result (Average of 3 trials)	Specs	JUDGMENT
DI Water	0.035	< 0.070	PASS
IPA	0.004	< 0.006	PASS

VI. Test for Specific Extractable Ions (μ g/g or ppm) - Standard Extraction Method (IEST-RP-C004.3 Section 7.2.2.18)

Extractable Ions	Result (Average of 3 trials)	Specs	JUDGMENT
Sodium (Na ⁺)	20.32	< 50.0	PASS
Potassium (K ⁺)	10.63	< 30.0	PASS
Calcium (Ca ²⁺)	13.18	< 30.0	PASS
Chloride (Cl ⁻)	10.42	< 30.0	PASS
Magnesium (Mg ²⁺)	11.24	< 50.0	PASS

The packaging materials, method & environment of this finished product are in accordance with Teknipure specifications.

Manufactured By : Production Technician: Zi Hong Wei

Inspected By: [Signature] 2018.11.27

Date of Manufacture: 2018/11/26

Teknipure/2150W.Broadway Rd./Suite 104/Mesa, AZ 85202/480-821-3182/WWW.teknipure.com