

Title	Teknipure C of C
Document Number	FM-8.2.5.1
Revision Date	February 20, 2022

# Certificate of Conformance

**Wiper Code: TZ1PCS1B-99**

**LOT No: 527166-241122**

Basis Weight (g/m <sup>2</sup> ) ASTM D3776		Thickness (mm) ASTM D1777-96		Orbital Shake Test for Particles IEST-RP-CC004.3 Section 6.1.4 (x10 <sup>6</sup> particle/m <sup>2</sup> )			Fiber/Particles IEST-RP-CC004.3 Section 6.2.2 (particle/m <sup>2</sup> )			Test For Sorbency Capacity & Rate IEST-RP-C004.3 Section 8.1						
Average	Lower SPEC Limit	Upper SPEC Limit	Thick-ness Average	Thick-ness Lower	Thick-ness Upper	>0.50 µm Average	>0.50 µm SPEC	>5.0 µm Average	>5.0 µm SPEC	>100 µm Average	>100 µm SPEC	Extrinsic Capacity (ml/m <sup>2</sup> ) Average	Intrinsic Capacity (ml/g) Average	Intrinsic Capacity SPEC	Sorbent Rate (Sec.) Average	Sorbent Rate (Sec.) SPEC
68	63	68	0.25	0.20	0.30	47.2	< 80.0	0.16	< 0.26	161	< 300	334	4.5	> 4.0	< 1	< 1

Extractables - Short Term Extraction (g/m <sup>2</sup> ) IEST-RP-CC004.3 Section 7.1.2		Specific Extractable Ions (ug/g or ppm) IEST-RP-C004.3 Section 7.2.2.1B											
DI Water Average	DI Water SPEC	IPA Average	IPA SPEC	Sodium (Na <sup>+</sup> ) Average	Sodium (Na <sup>+</sup> ) SPEC	Potassium (K <sup>+</sup> ) Average	Potassium (K <sup>+</sup> ) SPEC	Calcium (Ca <sup>2+</sup> ) Average	Calcium (Ca <sup>2+</sup> ) SPEC	Chloride (Cl <sup>-</sup> ) Average	Chloride (Cl <sup>-</sup> ) SPEC	Magnesium (Mg <sup>2+</sup> ) Average	Magnesium (Mg <sup>2+</sup> ) SPEC
0.042	< 0.07	0.005	< 0.006	17.2	< 50.0	9.5	< 30.0	8.3	< 30.0	11.6	< 30.0	12.8	< 50

**Product Cleanliness Specifications**  
The raw materials used in the production of this finished product are compliant with IEST-RP-CC004.3 and in accordance with Teknipure specifications.

**Product Packaging Specifications**  
The packaging materials, method & environment of this finished product are in accordance with Teknipure specifications.

**Shelf Life**  
Sterile product is marked with an expiration date three (3) years from date of manufacture.

Tested By: (Production Technician) Bill  
 Inspected By: (Quality) [Signature]  
 Name (printed): Bill  
 Date of Manufacture: Nov. 24, 2022



Innovative  
Contamination  
Solutions

Toll Free: 844.309.2376    Quality@Teknipure.com    www.Teknipure.com