

| | |
|-----------------|-------------------|
| Title | Teknipure C of C |
| Document Number | FM-8.2.5.1 |
| Revision Date | February 20, 2022 |

Certificate of Conformance

Wiper Code: TZ1PCS2PB-99

LOT No: 527238-250423

| Basis Weight (g/m ²) ASTM D3776 | | Thickness (mm) ASTM D1777-96 | | Orbital Shake Test for Particles IEST-RP-CC004.3 Section 6.1.4 (x10 ⁶ particle/m ²) | | | Fiber/Particles IEST-RP-CC004.3 Section 6.2.2 (particle/m ²) | | | Test For Sorbency Capacity & Rate IEST-RP-C004.3 Section 8.1 | | | | | | | |
|--|------------------------|---------------------------------|---------------------------|--|-------------------------|------------------------|---|-----------------------|--------------------|---|--------------------|----------------------------------|-------------------------------|----------------------------------|-------------------------------|------------------------------|---------------------------|
| Average | Lower SPEC Limit | Upper SPEC Limit | Thick- ness Average | Thick- ness Lower | Thick- ness Upper | >0.50 µm Average | >0.50 µm SPEC | >5.0 µm Average | >5.0 µm SPEC | >100 µm Average | >100 µm SPEC | Extrinsic Capacity Average | Extrinsic Capacity SPEC | Intrinsic Capacity Average | Intrinsic Capacity SPEC | Sorbitive Rate Average | Sorbitive Rate SPEC |
| 68 | 63 | 73 | 0.25 | 0.20 | 0.30 | 42.7 | 80.0 | 0.23 | 0.26 | 119 | 300 | 338 | 280 | 4.3 | 4.0 | < 1 | < 1 |

| Extractables - Short Term Extraction (g/m ²) IEST-RP-CC004.3 Section 7.1.2 | | | | Specific Extractable Ions (ug/g or ppm) IEST-RP-C004.3 Section 7.2.2.1B | | | | | | | | | |
|--|---------------------|----------------|-------------|--|--------------------------------------|---|--|---|--|---|--|---|--|
| DI Water Average | DI Water SPEC | IPA Average | IPA SPEC | Sodium (Na ⁺) Average | Sodium (Na ⁺) SPEC | Potassium (K ⁺) Average | Potassium (K ⁺) SPEC | Calcium (Ca ²⁺) Average | Calcium (Ca ²⁺) SPEC | Chloride (Cl ⁻) Average | Chloride (Cl ⁻) SPEC | Magnesium (Mg ²⁺) Average | Magnesium (Mg ²⁺) SPEC |
| 0.039 | < 0.07 | 0.0048 | < 0.003 | 15.3 | 50.0 | 13.7 | < 30.0 | 15.3 | < 30.0 | 12.7 | < 30.0 | 15.1 | < 50.0 |

Product Cleanliness Specifications

The raw materials used in the production of this finished product are compliant with IEST-RP-CC004.3 and in accordance with Teknipure specifications.

Product Packaging Specifications

The packaging materials, method & environment of this finished product are in accordance with Teknipure specifications.

Shelf Life

Sterile product is marked with an expiration date three (3) years from date of manufacture.

Tested By: (Production Technician) Bjll
 Inspected By: (Quality) Mow
 Name (printed): Bjll
 Date of Manufacture: Apr. 25, 2023



Innovative
Contamination
Solutions

Toll Free: 844.309.2376

Quality@Teknipure.com

www.Teknipure.com