

Title	Teknipure C of C
Document Number	FM-8.2.5.2
Revision Date	October 18, 2023

## Certificate of Conformance

**Wiper Code: TZ112S-25**

**LOT No: 527303-140923**

Basis Weight (g/m <sup>2</sup> ) ASTM D3776		Thickness (mm) ASTM D1777-96		Orbital Shake Test for Particles IEST-RP-CC004.3 Section 6.1.4 (x10 <sup>6</sup> particle/m <sup>2</sup> )		Fiber/Particles IEST-RP-CC004.3 Section 6.2.2 (particle/m <sup>2</sup> )		Test For Sorbency Capacity & Rate IEST-RP-C004.3 Section 8.1									
Average	Lower SPEC Limit	Upper SPEC Limit	Thickness Average	Thickness Lower	Thickness Upper	>0.50 µm Average	>0.50 µm SPEC	>5.0 µm Average	>5.0 µm SPEC	>100 µm Average	>100 µm SPEC	Extrinsic Capacity (ml/m <sup>2</sup> ) Average	Extrinsic Capacity SPEC	Intrinsic Capacity (ml/g) Average	Intrinsic Capacity SPEC	Sorbive Rate (Sec.) Average	Sorbive Rate (Sec.) SPEC
68	63	73	0.25	0.20	0.30	45.7	< 80.0	0.22	< 0.26	147	< 300	337	>280	4.5	> 4.0	< 1	< 1

Extractables - Short Term Extraction (g/m <sup>2</sup> ) IEST-RP-CC004.3 Section 7.1.2				Specific Extractable Ions (µg/g or ppm) IEST-RP-C004.3 Section 7.2.2.1B											
DI Water Average	DI Water SPEC	IPA Average	IPA SPEC	Sodium (Na <sup>+</sup> ) Average	Sodium (Na <sup>+</sup> ) SPEC	Potassium (K <sup>+</sup> ) Average	Potassium (K <sup>+</sup> ) SPEC	Calcium (Ca <sup>2+</sup> ) Average	Calcium (Ca <sup>2+</sup> ) SPEC	Calcium (Ca <sup>2+</sup> ) SPEC	Chloride (Cl <sup>-</sup> ) Average	Chloride (Cl <sup>-</sup> ) SPEC	Chloride (Cl <sup>-</sup> ) SPEC	Magnesium (Mg <sup>2+</sup> ) Average	Magnesium (Mg <sup>2+</sup> ) SPEC
0.032	< 0.07	0.004	< 0.006	12.5	< 60.0	11.3	< 30.0	8.5	< 30.0	< 30.0	11.5	< 30.0	< 30.0	12.7	< 50

**Product Cleanliness Specifications**  
The raw materials used in the production of this finished product are compliant with IEST-RP-CC004.3 and in accordance with Teknipure specifications.

**Product Packaging Specifications**  
The packaging materials, method & environment of this finished product are in accordance with Teknipure specifications.

**Shelf Life**  
Sterile product is marked with an expiration date three (3) years from date of manufacture.

Tested By: (Production Technician) B111  
 Inspected By: (Quality) MWL  
 Name (printed): B111  
 Date of Manufacture: Sep. 14, 2023

