Title	Teknipure C of C				
Document Number	FM-8.2.5.3				
Revision Date	November 15, 2023				

Certificate of Conformance

LOT No: 527365-091223 Wiper Code: TC2PU3-99

	s Weight (g ASTM D377			ickness (m TM D1777-	-		RP-CC004	est for Part .3 Section 6 rticle/m²)		IEST-RP- Sectio	Particles -CC004.3 n 6.2.2 cle/m²)	Test For Sorbency Capacity & Rate IEST-RP-C004.3 Section 8.1							
Average	Lower SPEC Limit	Upper SPEC Limit	Thick- ness Average	Thick- ness Lower	Thick- ness Upper	>0.50 µm Average	>0.50 µm SPEC	>5.0 µm Average	>5.0 µm SPEC	>100 µm Average	>100 µm SPEC	Extrinsic Capacity (ml/m²) Average	Extrinsic Capacity SPEC	Intrinsic Capacity (ml/g) Average	Intrinsic Capacity SPEC	Sorptive Rate (Sec.) Average	Sorptive Rate (Sec.) SPEC		
220	215	225	0.9	0.87	0.93	4.6	<6.0	0.14	<0.3	337	<350	727	>700	3.30	>3.15	<1	<1		

	(g/	ort Term Ext m²) l.3 Section 7		Specific Extractable lons (ug/g or ppm) IEST-RP-C004.3 Section 7.2.2.1B									
DI Water Average	DI Water SPEC	IPA Average	IPA SPEC	Sodium (Na ⁺) Average	Sodium (Na ⁺) SPEC	Potassium (K ⁺) Average	Potassium (K ⁺) SPEC	Calcium (Ca ²⁺) Average	Calcium (Ca ²⁺) SPEC	Chloride (Cl') Average	Chloride (Cl') SPEC	Magnesium (Mg ²⁺) Average	Magnesium (Mg ²⁺) SPEC
0.005	< 0.01	0.035	< 0.05	0.14	< 0.2	0.07	< 0.1	0.14	< 0.3	0.16	< 0.2	0.10	< 0.2

Product Cleanliness Specifications

The raw materials used in the production of this finished product are compliant with IEST-RP-CC004.3 and in accordance with Teknipure specifications.

Product Packaging Specifications

The packaging materials, method & environment of this finished product are in accordance with Teknipure specifications.

Tested By: (Production Technician)	张志馨
Inspected By: (Quality)	王雪艳
Name (printed):	王雪艳
Data of Manufactures Dec 0 2022	_
Date of Manufacture: <u>Dec.9.2023</u>	





