Title	Teknipure C of C
Document Number	FM-8.2.5.3
Revision Date	November 15, 2023

## **Certificate of Conformance**

## Wiper Code: TC2PU3-99

LOT No: 527431-060524

	s Weight (g ASTM D377		Thickness (mm) ASTM D1777-96			Orbital Shake Test for Particles IEST-RP-CC004.3 Section 6.1.4 (x10 <sup>6</sup> particle/m <sup>2</sup> )			Fiber/Particles IEST-RP-CC004.3 Section 6.2.2 (particle/m <sup>2</sup> )		Test For Sorbency Capacity & Rate IEST-RP-C004.3 Section 8.1						
Average	Lower SPEC Limit	Upper SPEC Limit	Thick- ness Average	Thick- ness Lower	Thick- ness Upper	>0.50 µm Average	>0.50 μm SPEC	>5.0 μm Average	>5.0 μm SPEC	>100 µm Average	>100 µm SPEC	Extrinsic Capacity (ml/m <sup>2</sup> ) Average	Extrinsic Capacity SPEC	Intrinsic Capacity (ml/g) Average	Intrinsic Capacity SPEC	Sorptive Rate (Sec.) Average	Sorptive Rate (Sec.) SPEC
220	215	225	0.9	0.87	0.93	4.4	<6.0	0.18	<0.3	342	<350	720	>700	3.32	>3.15	<1	<1

Extractables - Short Term Extraction (g/m <sup>2</sup> ) IEST-RP-CC004.3 Section 7.1.2				Specific Extractable lons (ug/g or ppm) IEST-RP-C004.3 Section 7.2.2.1B									
DI Water Average	DI Water SPEC	IPA Average	IPA SPEC	Sodium (Na⁺) Average	Sodium (Na⁺) SPEC	Potassium (K <sup>*</sup> ) Average	Potassium (K <sup>*</sup> ) SPEC	Calcium (Ca <sup>2+</sup> ) Average	Calcium (Ca <sup>2+</sup> ) SPEC	Chloride (Cl <sup>-</sup> ) Average	Chloride (Cl <sup>°</sup> ) SPEC	Magnesium (Mg <sup>2+</sup> ) Average	Magnesium (Mg <sup>2+</sup> ) SPEC
0.006	<0.01	0.031	< 0.05	0.11	<0.2	0.07	<0.1	0.16	<0.3	0.14	<0.2	0.16	<0.2

## Product Cleanliness Specifications

The raw materials used in the production of this finished product are compliant with IEST-RP-CC004.3 and in accordance with Teknipure specifications.

## Product Packaging Specifications

The packaging materials, method & environment of this finished product are in accordance with Teknipure specifications.

Tested By: (Production Technician)	张志豢
Inspected By: (Quality)	王雪艳
Name (printed):	王雪艳
Date of Manufacture: <u>May.6.2024</u>	

