

CERTIFICATE OF IRRADIATION



Certificate No.: GM2023122217

Contract No.: GM 23120586

R/SZJPY-7219 AD/00

Customer : Teknipure, LLC

Article Description	Carton Size(cm)	Article Lot No.	Carton Qty(cs)	G.W(kgs)
STERILE POLY/CEL WIPER TZ112S	61 * 58 * 50	527364	* 138 (12000bags) *	4311



THE PRODUCTS HAVE BEEN IRRADIATED BY GAMMARAY.

Irradiation Number: P2312174B

Irradiation Lot Finish Date: 2023/12/22

Minimum Specified Dose: 26.0 kGy Minimum Inspection Area Dose: 26.8 kGy

Maximum Specified Dose: 41.0 kGy Maximum Inspection Area Dose: 33.9 kGy

Inspector:  2023.12.22	SHENZHEN JPY ION-TECH. CO., LTD. GUANGMING BRANCH
Approved: 	
Date: 2023.12.22	