

Document Title	Teknipure C of A
Document Number	FM-8.2.4-4
Revision Date	February 12, 2024

Certificate of Analysis

Wiper Part No: TS2PUI100-44

Lot No: 417463

Component Traceability:

Component	PN#	LOT #
Wiper	TC2PUI-44	527385
Fluid	2-Propanol; VLSI/USP Purity	2AI240225B

I. Saturation & Fluid Specifications

	DIW Resistivity (M. Ohm)	DIW TOC (ppb)	Total Fluid Volume/Bag (ml)	IPA Fluid Volume/Bag (ml)	IPA Fluid %	DIW Fluid Volume/Bag (ml)	DIW Fluid %
Specification Limit	>17.8	<10	500	500	100	0	0

DIW – Deionized Water

II. Product Cleanliness Specifications

The raw materials used in the production of this finished product are compliant per IEST-RP-CC004.3 and in accordance with Teknipure upper limit specifications. A Certificate of Analysis for this lot of wipers can be obtained using the Component Wiper Lot# found in the first table on this page at <https://teknipure.com/certificate-of-conformance-search/>

III. Product Packaging Specifications

The packaging materials, method & environment of this finished product are in accordance with Teknipure specifications.

Manufactured By: (Production Technician) *E. V. [Signature]*
 Inspected By: (Quality) *Manuela [Signature]*
 Name (printed): Manuela [Signature]
 Date of Manufacture: 07/22/2024

