

| | |
|-----------------|-------------------|
| Title | Teknipure C of C |
| Document Number | FM-8.2.5.3 |
| Revision Date | November 15, 2023 |

Certificate of Conformance

Wiper Code: TC2P99S

LOT No: 527444-120624

| Basis Weight (g/m ²) ASTM D3776 | | | Thickness (mm) ASTM D1777-96 | | | Orbital Shake Test for Particles IEST-RP-CC004.3 Section 6.1.4 (x10 ⁶ particle/m ²) | | | | Fiber/Particles IEST-RP-CC004.3 Section 6.2.2 (particle/m ²) | | Test For Sorbency Capacity & Rate IEST-RP-C004.3 Section 8.1 | | | | | |
|--|------------------------|------------------------|---------------------------------|-------------------------|-------------------------|--|---------------------|--------------------|--------------------|---|--------------------|---|-------------------------------|--|-------------------------------|---------------------------------------|------------------------------------|
| Average | Lower SPEC Limit | Upper SPEC Limit | Thick- ness Average | Thick- ness Lower | Thick- ness Upper | >0.50 µm Average | >0.50 µm SPEC | >5.0 µm Average | >5.0 µm SPEC | >100 µm Average | >100 µm SPEC | Extrinsic Capacity (ml/m ²) Average | Extrinsic Capacity SPEC | Intrinsic Capacity (ml/g) Average | Intrinsic Capacity SPEC | Sorptive Rate (Sec.) Average | Sorptive Rate (Sec.) SPEC |
| 140 | 135 | 145 | 0.45 | 0.40 | 0.50 | 4.1 | < 5.0 | 0.4 | < 0.5 | 98 | <4 00 | 416 | > 400 | 2.98 | >2.8 | < 1 | < 1 |

| Extractables - Short Term Extraction (g/m ²) IEST-RP-CC004.3 Section 7.1.2 | | | | Specific Extractable Ions (ug/g or ppm) IEST-RP-C004.3 Section 7.2.2.1B | | | | | | | | | | |
|--|---------------------|----------------|-------------|--|--------------------------------------|---|--|---|--|---|--|---|---------------------------------------|--|
| DI Water Average | DI Water SPEC | IPA Average | IPA SPEC | Sodium (Na ⁺) Average | Sodium (Na ⁺) SPEC | Potassium (K ⁺) Average | Potassium (K ⁺) SPEC | Calcium (Ca ²⁺) Average | Calcium (Ca ²⁺) SPEC | Chloride (Cl ⁻) Average | Chloride (Cl ⁻) SPEC | Magnesium (Mg ²⁺) Average | Magnesium (Mg ²⁺) SPEC | |
| 0.0043 | < 0.01 | 0.021 | < 0.04 | 0.13 | < 0.20 | 0.07 | < 0.10 | 0.08 | < 0.20 | 0.07 | < 0.10 | 0.08 | < 0.10 | |

Product Cleanliness Specifications

The raw materials used in the production of this finished product are compliant with IEST-RP-CC004.3 and in accordance with Teknipure specifications.

Product Packaging Specifications

The packaging materials, method & environment of this finished product are in accordance with Teknipure specifications.

Tested By: (Production Technician) Bill
 Inspected By: (Quality) Mona
 Name (printed): Bill
 Date of Manufacture: Jun. 12, 2024

