

Title	Teknipure C of C
Document Number	FM-8.2.5.1
Revision Date	February 20, 2022

## Certificate of Conformance

Wiper Code: TZ1PCSS2PB-99

LOT No: 527315-25092301

Basis Weight (g/m <sup>2</sup> )		Thickness (mm)		Orbital Shake Test for Particles		Fiber/Particles		Test For Sorbency Capacity & Rate									
ASTM D3776		ASTM D1777-96		IEST-RP-CC004.3 Section 6.1.4 (x10 <sup>6</sup> particle/m <sup>2</sup> )		IEST-RP-CC004.3 Section 6.2.2 (particle/m <sup>3</sup> )		IEST-RP-C004.3 Section 8.1									
Average	Lower SPEC Limit	Upper SPEC Limit	Thickness Average	Thickness Lower	Thickness Upper	>0.50 µm Average	>0.50 µm SPEC	>5.0 µm Average	>5.0 µm SPEC	>100 µm Average	>100 µm SPEC	Extrinsic Capacity Average (ml/m <sup>2</sup> )	Extrinsic Capacity SPEC	Intrinsic Capacity Average (ml/g)	Intrinsic Capacity SPEC	Sorbent Rate Average (Sec.)	Sorbent Rate SPEC
68	63	73	0.25	0.20	0.30	47.2	80.0	0.20	0.26	112	300	338	280	4.5	4.0	< 1	< 1

Extractables - Short Term Extraction		Specific Extractable Ions (µg/g or ppm)											
IEST-RP-CC004.3 Section 7.1.2		IEST-RP-C004.3 Section 7.2.2.1B											
DI Water Average	DI Water SPEC	IPA Average	IPA SPEC	Sodium (Na <sup>+</sup> ) Average	Sodium (Na <sup>+</sup> ) SPEC	Potassium (K <sup>+</sup> ) Average	Potassium (K <sup>+</sup> ) SPEC	Calcium (Ca <sup>2+</sup> ) Average	Calcium (Ca <sup>2+</sup> ) SPEC	Chloride (Cl <sup>-</sup> ) Average	Chloride (Cl <sup>-</sup> ) SPEC	Magnesium (Mg <sup>2+</sup> ) Average	Magnesium (Mg <sup>2+</sup> ) SPEC
0.038	< 0.07	0.0044	< 0.006	12.5	50.0	10.2	< 30.0	15.2	< 30.0	15.3	< 30.0	15.7	< 50.0

**Product Cleanliness Specifications**

The raw materials used in the production of this finished product are compliant with IEST-RP-CC004.3 and in accordance with Teknipure specifications.

**Product Packaging Specifications**

The packaging materials, method & environment of this finished product are in accordance with Teknipure specifications.

**Shelf Life**

Sterile product is marked with an expiration date three (3) years from date of manufacture.

Tested By: (Production Technician)	Bill
Inspected By: (Quality)	Matt
Name (printed):	Bill
Date of Manufacture:	Sep. 25, 2023



Innovative Contamination Solutions

Toll Free: 844.309.2376 Quality@Teknipure.com www.Teknipure.com