

	Document Title:	ISO C of C
	Document Number:	FM-8.2.4-7
	Approved By:	Eric Hilkman
	Revision Date:	Twelve-27-2021

Certificate of Conformance

Wiper Code:ISO-MRC258S

Lot No:527611

I. Physical Property Variables

Property	Target	Lower	Upper	Units	Test Method
Individual weight	24	23	23	g	Weight Balance
Length	190	192	192	mm	Ruler
Width	65	66	66	mm	Ruler

II. Test for Particles (particles/cm²)- Orbital Shake Test(IEST-RP-CC004.3 Section 6.1.4)

Particle size	Result (Average of 3 trials)	Specs	JUDGMENT
> 0.5 um	1995	<4000	PASS

III. Test for Sorbency Capacity and Rate- (IEST-RP-C004.3 Section 8.1)Note: Unless specified the Default Test Solution is 01 Water

Absorbency	Result (Average of 3 trials)	Specs	JUDGMENT
Extrinsic Capacity (mL/pcs)	66	>60	PASS
Sorptive Rate (s)	0.48	<1	PASS

IV. Test for Extractable Matters (mg/g)• Short Term Extraction

Test Solution	Result (Average of 3 trials)	Specs	JUDGMENT
DI Water	0.0112	<1.0	PASS
DI IPA	0.2548	<1.0	PASS

The packaging materials,method & environment of this finished product are in accordance with Teknipure specifications.

Tested By: Liang Jia En

Laboratory Manager: Xu Ying

Date of Manufacture: 02-25-2025