

CERTIFICATE OF IRRADIATION



Certificate No. GM2025072534


Contract No. : GM25070723

R-SZJPY-7219 AE-00

Customer: Teknipro, LLC

Product	Carton Size(cm)	Product Lot No.	Qty(ctns)	G.W.(kg)
STERILE CLEANROOM WIPER TC2P99S P1-LP9D-ST	27*27*30	Lot No: 212502 Manufacture Date: 07-2025 Expiration Date: 06-2028 20pcs/bag, 25bags/case	*55*	220

THE PRODUCTS HAVE BEEN IRRADIATED BY GAMMARAY.

Irradiation Code No. :	P2507225B
Irradiation lot finish date:	2025/07/25
Required minimum absorbed dose: 25.0 kGy	Irradiation minimum absorbed dose: 28.1 kGy
Required maximum absorbed dose: 45.0 kGy	Irradiation maximum absorbed dose: 37.7 kGy
Inspector/Date:  2025.7.26	SHENZHEN JPY ION-TECH. CO., LTD.
Approver/Date:  2025.7.26	