


CERTIFICATE OF IRRADIATION



Certificate No.: GM2025071915

Contract No.: GM25070514

R/SZJPY-7219 AD/02

Customer: Teknipure LLC				
Article description	Carton Size(cm)	Article Lot No.	Carton Qty(cs)	G.W(kgs)
STERILE POLY/CEL WIP ER TM-QP258S	59*59*49	527684	*6 (396bags)	* 172.2

THE PRODUCTS HAVE BEEN IRRADIATED BY GAMMARAY.

Irradiation No.	P2507160B
Irradiation finish date	2025/07/19
Required minimum absorbed dose: 26.0 kGy	Irradiation minimum absorbed dose: 29.0 kGy
Required maximum absorbed dose: 41.0 kGy	Irradiation maximum absorbed dose: 40.2 kGy
Inspector/Date: 2025.7.19	SHENZHEN JPY ION-TECH. CO., LTD. GUANGMING BRANCH
Approved/Date: 2025.7.19	