

Certificate of Processing

STERIGENICS 344 Bonnie Circle Corona CA 92878
TEL 951 340-0700 FAX 951 340-1990 www.sterigenics.com

R55480102

06/25/2024 15:22:33 GMT
Page 1 of 1

Customer Name:	Teknipure, a Division of Enefco USA	Processing Facility:	Corona	Work Order #	3574316
P.O.#	527486			Sales Order #	3578213
	25-40 kGy	TKN_25-40, Gamma Treatment	Irradiation Date/Time:	06/23/2024 07:16:00 GMT	
			Irradiation Cell:	CELL A	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	41	CA	TS1MPI70-911S		417412	NA
	Dose Map		104_19264	HI: TE1 LO: 1C9		
	41	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	40.0 kGy	27.3 KGY	Pass	MCHAMU	06/23/2024 22:16:19 GMT
		Reason Code Test				Matias Chamu	
450.00	Maximum Dose	25.0 kGy	40.0 kGy	33.3 KGY	Pass	MCHAMU	06/23/2024 22:16:29 GMT
		Reason Code Test				Matias Chamu	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: RAYLENE TOVAR
Reason: Work Order Completions

Date: 06/25/2024 15:20:51 GMT