

Certificate of Processing

STERIGENICS 344 Bonnie Circle Corona CA 92878
 TEL 951 340-0700 FAX 951 340-1990 www.sterigenics.com

R55480102

10/03/2025 15:49:13 GMT
 Page 1 of 1

Customer Name:	Teknipure, a Division of Enefc USA	Processing Facility:	Corona	Work Order #	3844208
P.O.#	527737			Sales Order #	3908173
	25-45 kGy	TKN_25-45, Gamma Treatment		Irradiation Date/Time:	10/02/2025 16:05:00 GMT
				Irradiation Cell:	CELL A

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	30	CA	TS1PCI70-99S/50		382501	n/a
	Dose Map		104_21170-21171-21172	TOTE TAG AFMIN0.691 AFMAX1.061		
	30	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	26.1 KGY	Pass	JGARCIA	10/03/2025 05:53:36 GMT
		Reason Code Test				Jose Garcia	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	40.5 KGY	Pass	JGARCIA	10/03/2025 05:53:49 GMT
		Reason Code Test				Jose Garcia	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: RAYLENE TOVAR
 Reason: Work Order Completions

Date: 10/03/2025 15:48:22 GMT