

CERTIFICATE OF IRRADIATION

Certificate No. : GM2025111310

Contract No. : GM25110352

R-SZJPY-7219 AE-00

Teknipure

Product	Carton Size(cm)	Product Lot No.	Qty(ctns)	G.W.(kg)
STERILE CLEANROOM WIPER TC2P12S P1-LP12D-ST	34*34*27	Lot No:452502 Manufacture Date:11-2025 Expiration Date:11-2028 25pcs/bag, 20bags/case	*100*	715

THE PRODUCTS HAVE BEEN IRRADIATED BY GAMMARAY.

Irradiation Code No. :	P2511085B
Irradiation lot finish date:	2025/11/13
Required minimum absorbed dose: 25.0 kGy	Irradiation minimum absorbed dose: 26.2 kGy
Required maximum absorbed dose: 45.0 kGy	Irradiation maximum absorbed dose: 35.2 kGy
Inspector/Date: 2025/11/13	 SHENZHEN JPY ION-TECH. CO., LTD.
Approver/Date: John Condon 2025.11.13	