

CERTIFICATE OF IRRADIATION



Certificate No GM2025102108

Contract No. : GM25100390

R-SZJPY-7219 AE-00

Customer: Teknipure, LLC				
Product	Carton Size(cm)	Product Lot No.	Qty(ctns)	G.W.(kg)
STERILE CLEANROOM WIPER TC2P99S	27*27*30	Lot No:402502 Manufacture Date:09-2025 Expiration Date:09-2028 20pcs/bag, 25bags/case	*50*	200
THE PRODUCTS HAVE BEEN IRRADIATED BY GAMMARAY.				
Irradiation Code No. :		P2510147B		
Irradiation lot finish date:		2025/10/21		
Required minimum absorbed dose: 25.0 kGy		Irradiation minimum absorbed dose: 28.9 kGy		
Required maximum absorbed dose: 45.0 kGy		Irradiation maximum absorbed dose: 33.4 kGy		
Inspector/Date:  2025.10.21		SHENZHEN JPY ION-TECH. CO., LTD.		
Approver/Date:  2025.10.21				