

**CERTIFICATE OF IRRADIATION**

Certificate No.: GM2025111338

Contract No.: GM25110352

R-SZJPY-7219 AE-00

Product	Carton Size(cm)	Product Lot No.	Qty(ctns)	G.W.(kg)
STERILE NONWOVEN CLEAN WIPER TZ112S-25 SK3-M612-ST-25	34*34*19	Lot No:442502 Manufacture Date: 11-2025 Expiration Date :10-2028 25 pcs /inner bag 2 inner bags/ Outer bag 10 Outer bags / case	*30*	146.1

**THE PRODUCTS HAVE BEEN IRRADIATED BY GAMMARAY.**

Irradiation Code No.:	P2511089B
Irradiation lot finish date:	2025/11/13
Required minimum absorbed dose: 29.5 kGy	Irradiation minimum absorbed dose: 33.9 kGy
Required maximum absorbed dose: 50.0 kGy	Irradiation maximum absorbed dose: 44.5 kGy
Inspector/Date:	
Approver/Date: <i>John Condon</i> 2025.11.14	SHENZHEN JPY ION-TECH. CO., LTD.