

CERTIFICATE OF IRRADIATION



Certificate No. GM2025092829

Contract No. : GM25090856

R-SZJPY-7219 AE-00

Customer: Teknipure, LLC				
Product	Carton Size(cm)	产品批号 Product Lot No.	Qty(ctns)	G.W.(kg)
STERILE NONWOVEN CLEAN WIPER TZ112S-25	34*34*19	Lot No:262502 Manufacture Date:07-2025 Expiration Date :06-2028 25 pcs /bag 20 bags / case	*20*	119.6

THE PRODUCTS HAVE BEEN IRRADIATED BY GAMMARAY.

Irradiation Code No. :	P2509288B
Irradiation lot finish date:	2025/09/28
Required minimum absorbed dose: 29.5 kGy	Irradiation minimum absorbed dose: 32.9 kGy
Required maximum absorbed dose: 50.0 kGy	Irradiation maximum absorbed dose: 40.5 kGy
Inspector/Date:  2025.10.3	SHENZHEN JPY ION-TECH. CO., LTD.
Approver/Date:  2025.10.3	